



## EU - Declaration of Conformity

<b>Manufacturer:</b>	GN Hearing A/S Lautrupbjerg 7 DK-2750 Ballerup Denmark
<b>Single Registration Number (SRN):</b>	DK-MF-000009536
<b>Conformity Assessment Procedure (MDR):</b>	Article 52(7) of the Regulation (EU) 2017/745
<b>Conformity Assessment Procedure (RED):</b>	Annex III of the Radio Equipment Directive (RED) 2014/53/EU
<b>Identification of the Device:</b>	<b>Category:</b> Accessory
	<b>Device model:</b> Multi Mic
	<b>Brand:</b> Refer to Appendix A
	<b>Model:</b> Refer to <i>Description</i> in Appendix A

**Classification of the Device (MDR):** Class I, Rule 1 and 13

### Common specifications:

-

### Applied standards and normative standards:

RED: Health & Safety: IEC 62368-1:2018, EN IEC 62479:2010;  
EMC: ETSI EN 301 489-17 V3.2.6 (2023-06), ETSI EN 301 489-1 V2.2.3 (2019-11), IEC 62368-3:2017, EN 60601-1-2:2015+A1:2021;  
Spectrum: ETSI EN 300 328 V2.2.2 (2019-07);


GN Hearing A/S, the manufacturer, hereby declare, under our sole responsibility, that the below-mentioned devices comply with the relevant provisions of:

- REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL (MDR), as amended
- DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL (RED), as amended
- DIRECTIVE 2011/65/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL (RoHS), as amended

and, where applicable, their relevant transpositions into national laws of the EU Member States in which the below-mentioned devices are distributed.

**Place and Date: Ballerup, Denmark 2025-11-06**

Signed by:

  
D6EFFFFCB74DF41A...

**Karen Bryde Frisk**

Director of Regulatory Affairs

Corporate Quality

GN Hearing A/S



## **Appendix A**

The following listed models are compliant to the requirements stated on page 1.

### **Category:**

Accessory

### **Intended Purpose:**

The Multi Mic is intended to be a wireless remote microphone or connection to an audio device for wireless hearing aids made by GN independent of brand.

When used as a remote microphone, the Multi Mic is placed close to a person or a sound source and sends the voice or sound directly to the hearing aid.

When used as a connection to an audio device, the Multi Mic is connected to the audio device through cable, Telecoil or an FM adapter and the sound is sent directly to the hearing aid.

The Multi Mic including BLE functionality may also be used together with Approved Partner Products. Approved Partner Product refers to specific products released and marketed by 3rd party partners .

<b>Brand</b>	<b>Family Name</b>	<b>Product Name</b>	<b>Description</b>	<b>Item Number</b>	<b>Product Form Factor (PFF)</b>	<b>Basic UDI-DI</b>	<b>GMDN Code</b>	<b>EMDN Code</b>
GN Hearing	N/A	MULTI-MIC+	MULTI-MIC+,SM-3P	23093300	SM-3P	57082960000393N	57886	Y214580
GN Hearing	N/A	MULTI-MIC+	MULTI-MIC+,SM-3P,DEMO	23325800	SM-3P	57082960000393N	57886	Y214580